

6390 Joyce Drive Phone 303-940-0033
100 Fax 866-283-0269
Golden, CO 80403 www.wibby.com

Final Report - Water Supply Proficiency Testing

Study: WS1011

Opening Date: October 3, 2011 - Closing Date: November 17, 2011

Laboratory: Chemsolutions
7388 S. Revere Parkway #806
Centennial, CO 80112

Contact: John Graves
303-771-5570

EPA Lab ID: CO01058

EDB/DBCP/TCP (PT-EDBCP-WS)

Lot #: 9046-27

NELAC Code	Analyte	Analysis Date	Analyst	Method Code	Method Description	Units	Assigned Value	Result	Acceptance Limits	Evaluation
4570	1,2-Dibromo-3-chloropropane (DBCP)	11/10/2011	John Graves	10173009	EPA 8011 1994	µg/L	1.45	1.34	0.871 - 2.03	Acceptable
4585	Ethylene Dibromide (EDB)	11/10/2011	John Graves	10173009	EPA 8011 1994	µg/L	1.92	1.52	1.15 - 2.69	Acceptable
Additional State Specific Analytes										
5180	1,2,3-Trichloropropane					µg/L	1.79		1.08 - 2.51	NR

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.wibby.com

PT Study Score Report

EPA ID: CO01058
Laboratory: Chemsolutions
7388 S. Revere Parkway #806
Centennial, CO 80112

Study Number	Number of Reported Results	Number of Passing Results	Percent Passing
WS1011	2	2	100%

Report Definitions:

- Number of Reported Results*** The number of results reported which could be evaluated. Results receiving an evaluation of "NR" or "No Evaluation" are not included.
- Number of Passing Results*** The number of results reported receiving an evaluation of "Acceptable" or "Check for Error".
- Percent Passing*** The percentage of results reported receiving an evaluation of "Acceptable" or "Check for Error".